

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60607

Customer	Al - Ory Hospital Jahra			Equipment			
				Model No.	G-Runner 322		
				Serial No.	21000619		
Authorized Person				Manufacturer	Polema		
Customer Complaint				Location			
laser pulse not moving up, no laser on gun.				Contract / Order No.			
Services / Check Carried out				☐ Installation ☐ Warranty			
				☐ Repair ☐ On Charge			
☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐				☐ Maintenance ☐ Contract			
				☐ Upgrade ☐			
				Time Spent			
Remarks				Noted Fault in Interface 8V0-21000658 Replaced Interface with new interface model - 8V0-21000636 checked G Runner operations found OK			
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer		Authorized Person		Customer / User			
Arun-K		[Signature]					
ATCO Technology Co.							
Signature		Signature		Signature			
Date		Date		Date			