

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60673

Customer	Mot Hognif	Equipment	Dental Chair.
		Model No.	SS-KISS
		Serial No.	2112438
Authorized Person		Manufacturer	Simple & Smart
Customer Complaint		Location	Dental.
		Contract / Order No.	
	low suction n/w.	☐ Installation	☐ Warranty
		☐ Repair	☐ On Charge
		☐ Maintenance	☐ Contract
		☐ Upgrade	☐ _____
Services / Check Carried out		Time Spent	
☐ Inspection / Cleaning	☐ Optics Checks		
☐ Calibration / Performance	☐ Pneumatic Checks		
☐ Interface / Communication	☐ Hydraulic Checks		
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Found low suction hose blocked with patient materials - cleaned hose, unit is working O.K.		

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Signature Date	Authorized Person Signature Date	Customer / User Signature Date
 ATCO Technology Co. 18/7/24	 د. عبد الرحمن السعيد عبد الله طبيب أسنان مسجل أول الاستشارة الصناعية مستشفى السليمية 22019100056	