

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٢٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61988

Customer	Alor Hospital			Equipment	Shock wave therapy unit		
				Model No.	SHOCK MED		
				Serial No.	EM09311018		
Authorized Person				Manufacturer	EME		
Customer Complaint	Touch screen replacement			Location	Physiotherapy		
				Contract / Order No.			
Services / Check Carried out				☐ Installation	☐ Warranty		
☐ Inspection / Cleaning	☐ Optics Checks			☒ Repair	☐ On Charge		
☐ Calibration / Performance	☐ Pneumatic Checks			☐ Maintenance	☐ Contract		
☐ Interface / Communication	☐ Hydraulic Checks			☐ Upgrade	☐ _____		
☐ Safety Checks	☐ Operational Error			Time Spent			
☐ Electrical / Electronics	☐ Mechanical Checks						
☐ Accessories Check	☐ _____						
Remarks	we replaced the touch screen with new one. Tested and calibrated the touch screen. checked and found that the unit is working properly.						
No.	Spare Parts Used		Part No.	Qty.			
1.	Touch screen			1			
Service Engineer	Authorized Person		Customer / User				
Mixor/Mouse							
ATCO Technology Co.							
Signature	Signature		Signature				
Date	Date		Date				
27/08/2024							