

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218



شركة أتكو التكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)
فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)
ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت
Email: sales@atco-kuwait.com
www.atco-kuwait.com
س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63661

Customer	Aloft Hospital	Equipment	Laser machine
		Model No.	SP Dynamics
		Serial No.	16100425
Authorized Person		Manufacturer	Fotona
Customer Complaint		Location	4th floor
Error 421		Contract / Order No.	
Services / Check Carried out		☐ Installation ☐ Warranty	
☐ Inspection / Cleaning		☒ Repair	
☐ Calibration / Performance		☐ On Charge	
☐ Interface / Communication		☐ Maintenance	
☐ Safety Checks		☐ Contract	
☐ Electrical / Electronics		☐ Upgrade	
☐ Accessories Check		☐ _____	
		Time Spent	

Remarks We found that the Charger module is faulty. we fixed the unit with our stand by charger module. we are taking the machine charger module to our workshop. once it repaired we will return back the charger unit.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Nixon ATCO Technology Co. Signature Date 19/03/2025	Authorized Person Signature Date	Customer / User د. ايمن كويك استشاري أمراض جلدية وخصائية مستشفى العرف 140925 Signature Date
--	--	---