

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63799

Customer	Aper Hospital Jahra			Equipment	Derma Laser					
				Model No.	SP Dynamis					
				Serial No.	16100425					
Authorized Person				Manufacturer	Futura					
Customer Complaint				Location	4th FLR. DERMA					
				Contract / Order No.						
Services / Check Carried out				☐ Installation	☐ Warranty					
				☐ Repair	☐ On Charge					
				☐ Maintenance	☐ Contract					
				☐ Upgrade	☐ _____					
				Time Spent						
Remarks				Found the GUI and sys software needed to be updated for the installation of new honey comb adapter. Will send quotes for the software.						
No.	Spare Parts Used			Part No.	Qty.					
Service Engineer				Authorized Person		Customer / User				
Surith						د. ايمن توبيا نواف				
ATCO Technology Co.						استشاري اجهزة طبية وتناسلية				
Signature				Signature		Signature				
Date				Date		Date				
08/01/2025						140625				