

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 64284

Customer	Alorj Hospital		Equipment	Saline warmer	
	OT		Model No.		
			Serial No.		
Authorized Person			Manufacturer	Hagawi	
Customer Complaint	Digital thermometer not working		Location	Intensive OT	
			Contract / Order No.		
Services / Check Carried out			☐ Installation	☐ Warranty	
☐ Inspection / Cleaning	☐ Optics Checks		☒ Repair	☐ On Charge	
☐ Calibration / Performance	☐ Pneumatic Checks		☐ Maintenance	☐ Contract	
☐ Interface / Communication	☐ Hydraulic Checks		☐ Upgrade	☐ <u>EM</u>	
☐ Safety Checks	☐ Operational Error		Time Spent		
☐ Electrical / Electronics	☐ Mechanical Checks				
☐ Accessories Check	☐				
Remarks	Digital thermometer not working				
Replaced the battery thermometer working good					
All connections are tested OK					
Warmer Fixed working good.					
No.	Spare Parts Used		Part No.	Qty.	
1	Battery			1	
	won't work now				
Service Engineer		Authorized Person		Customer / User	
Jince Cherian		22/1/25			
ATCO Technology Co.					
Signature		Signature		Signature	
Date 22/01/2025		Date		Date	