

INCIDENT REPORT FORM

LOG #:

Forward Completed Form to Department Manager. One copy to Quality Management Department.

PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) Radiology
 Date of Occurrence: 25/5/25 Time of Occurrence: 25/5/25 Occurrence Location: _____
 If event involved employee - Staff Name: Angela - p.l. @ 11.30h Age: 31 Sex: fe Emp#: 4123
 If event involved patient - Patient Name: Emp no - 4123 Age: _____ Sex: _____ File#: _____
 If event involved Visitor - Visitor Name: STAFF NURSE Phone No.: _____
 Address: _____

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line	☐ Delays	☐ Burn
☐ Contrast Media Reaction	☐ Missing Files, Notes	☐ Chemical Spill
☐ Extravasate	☐ Performed on Wrong Site	☐ Electrical Shock
☐ Occluded	☐ Consent	☐ Fire
☐ Precipitate	☐ Patient Identification	☐ Gas
☐ Site Red/Swollen	☐ Wrong Patient	☐ Hazardous Waste
☐ Tubing Leak	☐ Wrong Procedure	☐ Radiation Exposure
☐ Expired IV Fluid	☐ Lost/Spoiled Specimen	☐ Splash
☐ Additive Related	☐ No Response	☐ Water Spill
☐ Transfusion Reaction	☐ Wrong File	☐ Needle prick injury
☐ Other _____	☐ Procedure Cancellation	☐ Other _____
	☐ Notes in Wrong File	
	☐ Inadequate/Inappropriate Entry	
	☐ No label/Addressograph	
	☐ Other _____	
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp	☐ Damage/Loss of Property	☐ Electrical Problem
☐ AMA/Absconded	☐ Kidnapping	☒ Not Working
☐ Dissatisfied with care	☐ Missing/Misplaced Drug Keys	☐ Improper Use
☐ Patient Behavior	☐ Prescription Alteration	☐ Mechanical Problem
☐ Family/Visitor Behavior	☐ Security Problem	☐ Operator Not Qualified
☐ Staff Behavior	☐ Suicide attempt	☐ Wrong Equipment
☐ Time In/Time Out Variation	☐ Other _____	☐ Missing Equipment
☐ Non-Compliance		☐ Not Available
☐ Patient Complaint		☐ Out of Stock
☐ Other _____		☒ Other <u>SPO2 probe</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by Name: Angela ID No. 4123 Dept. Radiology Position STAFF NURSE

IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication

Assessment/
Diagnosis

Physician Name:

Dept.:

ID No.:

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

In patient monitor, the spo₂ probe is not working.

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
I reported the incident to the nurse dept.		Faulty device referred to	
I advised the nurse to always take care		to medical for checking	
the handles, the machine in the department.		repair.	
Dept. Head: <u>Nursing</u>	Date: <u>27/5/25</u>	Quality Mgmt. Dept. <u>Charlton T.D/KW</u>	Date: <u> </u>
ID No. <u>1022</u>	Department: <u>NURSING</u>	Comments: <u> </u>	

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 07/05/25

Request by Anjuna

Department Radiology Location _____

SR. No.	Machine	Model	Complaint
<u>E0850060</u>	<u>cardiac monitor</u>	<u>Vitapa 2000</u>	<u>spo2 probe not working</u>

Check within Week

ASAP

Engineer Comments

End Comments

Anjuna

Request by

Checked by

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO : 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب. ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63673

Customer	Alert Hospital	Equipment	Patient monitor
		Model No.	Vitala 7000
		Serial No.	E 085 D0060
Authorized Person		Manufacturer	Tosmed
Customer Complaint		Location	X-Ray
	SPO2 not work	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☒ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ <u>Em</u>
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____	Time Spent	
Remarks	Service call attended. Found SPO2 probe defective - replaced SPO2 probe and confirmed unit is working OK.		
No.	Spare Parts Used	Part No.	Qty.
	SPO2 probe		1
Service Engineer	Authorized Person	Customer / User	
ATCO Technology Co.			
Signature	Signature	Signature	
Date	Date	Date	
25/5/25	27/5/25		