

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكوتكنولوجيه

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهره, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 65489

Customer	ALORF HOSPITAL	Equipment	Washer & Disinfection
	Jahra	Model No.	Unichan PL
		Serial No.	230506
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	CSSD
Unloading door not open		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐		☐ Installation	☐ Warranty
		☐ Repair	☐ On Charge
		☐ Maintenance	☐ Contract
		☐ Upgrade	☐
		Time Spent	

Remarks: Checked Bunnel unloading door not opened, Intermittently after the cycle, suspected problem with the door gasket and relay module. Replaced door gasket and relay module. One cycle fully completed. All connections are checked. System working good.

① Unloading door not open due to technical problems.

No.	Spare Parts Used	Part No.	Qty.
1	Door Seal		1
2	Complete relay module		1
	9/6/25 - TEST done.		
	23/6/25 - TEST done.		

Service Engineer ATCO service team ATCO Technology Co. Signature Date 25/06/25	Authorized Person Signature Date 28/6/25	Customer / User Signature Date
--	--	--------------------------------------