

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجي

تلفون : ٢٤٧٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٧٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء، ٠١٠٢٥ - الكويت


Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٢٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 56420

Customer	Alor Hospital Jahra		
Equipment	Gynae Laser		
Model No.	SP Dynamis		
Serial No.	19101628		
Authorized Person	Fibona		
Customer Complaint	Location		
	Contract / Order No.		
Services / Check Carried out	☐ Installation ☒ Warranty ☒ Repair ☐ On Charge ☐ Maintenance ☐ Contract ☐ Upgrade ☐ _____		
☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐ _____	Time Spent		
Remarks	Returned the A-Rumer SN: 21000619 with accessories taken. Found problem with the interface controller. Given strongly Replaced the interface controller under warranty. Checked and found it was not properly. the New interface controller - SN: 21000658		
No.	Spare Parts Used	Part No.	Qty.
	GRA - SN: 752		
	GRA - SN: 774		
	GRD - SN: 390		
	Ex. YAG shub - 191580		
	Nb. YAG shub - 139823		
Service Engineer	Authorized Person	Customer / User	
Signature 	Signature 	Signature	
Date 03/01/2024	Date	Date	