

INCIDENT REPORT FORM

LOG #: _____ **Forward Completed Form to Department Manager. One copy to Quality Management Department.**

PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____

Date of Occurrence: 11-5-2023 Time of Occurrence: 5 AM Occurrence Location: LAB

If event involved employee - Staff Name: Vinceha Kojan Age: 33 Sex: F Emp#: 2741

If event involved patient - Patient Name: _____ Age: _____ Sex: _____ File#: _____

If event involved Visitor - Visitor Name: _____ Phone No.: _____

Address: _____

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☒ Not Working ☐ Improper Use ☐ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>CENTRIFUGE</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by Name: MARILY CORPUZ ID No. 1645 Dept. LAB Position LAB TECH

IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication

**Assessment/
Diagnosis**

Physician Name: _____ Dept.: _____ ID No.: _____

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

The centrifuge lead is not closing.
 When we start the morning duty, about to
 run the control the Maclian Centrifuge
 not working (Low Problem).
 SN - 60150147 Model HERMLE

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
☐ **Moderate.** Increased length of Stay. Increased level of care.
☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
inform the Biomedical department		- Referred to Biomedical for repair - Use until cleared	
Dept. Head: <i>Dr. Asaf Sang</i>	Date: <i>11-5-2025</i>	Quality Mgmt. Dept. <i>Star late TB (R)</i>	Date:
ID No.	Department:	Comments:	

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 11-5-2025

Requested by MARLY

Department LAB Location Clinical Microscopy

SR. No.	Machine	Model	Complaint
60150147	Centrifuge	HERMLE	The lead is not closing

Check within Week

ASAP

Engineer Comments

End Comments

MARLY
Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكوتكنولوجيه

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجبراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63664

Customer	Alorj Hospital			Equipment	Centrifuge		
				Model No.	22064		
				Serial No.	As below		
Authorized Person				Manufacturer	Hermle		
Customer Complaint	Lid lock.			Location	Laboratory		
				Contract / Order No.			
Services / Check Carried out				☐ Installation	☐ Warranty		
☐ Inspection / Cleaning	☐ Optics Checks			☒ Repair	☐ On Charge		
☐ Calibration / Performance	☐ Pneumatic Checks			☐ Maintenance	☐ Contract		
☐ Interface / Communication	☐ Hydraulic Checks			☐ Upgrade	☐ _____		
☐ Safety Checks	☐ Operational Error			Time Spent			
☐ Electrical / Electronics	☐ Mechanical Checks						
☐ Accessories Check	☐ _____						
Remarks	We replaced the three centrifuge lid lock with new one. checked and confirmed that the three units are working properly.						
	Serial number: - 601900343, 601900026, 60150167						
No.	Spare Parts Used			Part No.	Qty.		
1-	lid lock				3		
Service Engineer	Authorized Person			Customer / User			
Signature	Signature			Signature			
Date	Date			Date			
18/05/2025				18-5-2025			
				د. ايات سامي محمد عبد الله			
				طبيب مسجل باثولوجي			
				مستشفى العرف			
				22023120560			