

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



log no - 2195

شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجاهراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61980

Customer	Aloof Hospital	Equipment	Electro Surgical Unit
		Model No.	Surgilec MC3
		Serial No.	208
Authorized Person		Manufacturer	Surgilec
Customer Complaint		Location	OT
	Surgical handpiece not detecting.	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		

Remarks Service call attended.

We inspect the unit and found that the connection for the surgical handpiece connector is broken. We repaired the connector and fixed it properly. Checked and confirmed that the unit is working properly.

No.	Spare Parts Used	Part No.	Qty.
	4 times returned for repair.		

Service Engineer Nidom ATCO Technology Co. Signature Date 18/02/2024	Authorized Person Signature Date 25/8/24	Customer / User Signature Date
--	--	--------------------------------------