

MARCH 9, 2025

1. ANESTHESIA MACHINE

SN # AMXH00507

PROBLEM: MONITOR NOT WORKING

REPLACEMENT OF POWER SUPPLY,
CALIBRATION AND CLEANING OF MONITOR
POWER SUPPLY.

60 KD

to PM CASH

to Mr. EMMANUEL

Handwritten signature

د. ماجد يحيى ريان
اختصاص في التخدير
مستشفى السرطان
133187

OK

Handwritten signature