

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63790

Customer	Ahraf Hospital	Equipment	Defibrillation
		Model No.	Cardio Aid 360
		Serial No.	20206122
Authorized Person		Manufacturer	Innomed
Customer Complaint		Location	Dental
Battery failure		Contract / Order No.	
Services / Check Carried out		☐ Installation ☐ Warranty	
☐ Inspection / Cleaning		☒ Repair	
☐ Calibration / Performance		☐ On Charge	
☐ Interface / Communication		☐ Maintenance	
☐ Safety Checks		☐ Contract	
☐ Electrical / Electronics		☐ Upgrade	
☐ Accessories Check		☐ _____	
		Time Spent	
Remarks			
Service call attended.			
We inspect the unit and found that the battery failure. we discharge the battery and cleaned the error. Now the unit is working properly.			
Checked and confirmed that the unit is working properly.			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer		Authorized Person	
Jins / Alson			
ATCO Technology Co.			
Signature		Signature	
Date 30/12/2024		Date	
		Customer / User	
		Signature	
		Date	