

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرام، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60665

Customer	Hasef Hospital			Equipment	ENT Unit.		
				Model No.	Futerenit II		
				Serial No.	111 63 2045-		
Authorized Person				Manufacturer	Entesmed.		
Customer Complaint				Location	Ent - 5th Flr.		
	Ethesmo Not Heating.			Contract / Order No.			
				☐ Installation	☐ Warranty		
				☐ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐ Em		
				Time Spent			
Services / Check Carried out							
☐ Inspection / Cleaning			☐ Optics Checks				
☐ Calibration / Performance			☐ Pneumatic Checks				
☐ Interface / Communication			☐ Hydraulic Checks				
☐ Safety Checks			☐ Operational Error				
☐ Electrical / Electronics			☐ Mechanical Checks				
☐ Accessories Check			☐				
Remarks	Replaced ethesmo mainboard, replaced pressure switch adjusted, tested and confirmed unit is working properly.						
No.	Spare Parts Used			Part No.	Qty.		
	ethesmo main board				1		
Service Engineer	Authorized Person			Customer / User			
Signature	Signature			Signature			
Date	Date			Date			
	12/6/24			د. ايضان سلافوف استشاري انف واد وخدمات مستشفى البشير 150952			