

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية Log no - 1722

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58948

Customer	Alof Hospital	Equipment	Sperm Analysis.
		Model No.	Lengthake XI PRO
		Serial No.	X10VL16 M 10007
Authorized Person		Manufacturer	Taba Biomaybio
Customer Complaint		Location	laboratory
Network connection issue		Contract / Order No.	

Services / Check Carried out

- | | |
|--|--|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ |

- | | |
|---|---------------------------------|
| ☐ Installation | ☐ Warranty |
| ☒ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ |

Time Spent

Remarks

we checked the unit and found that the fire wall is blocking the unit IP address with the help of hospital IT we resolve the firewall issue.

currently the unit is connected with the PC checked and confirmed that the unit is working properly

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

Nidom

ATCO Technology Co.

Signature

Date 26/03/2024

Authorized Person

Signature

Date

Customer / User

د. سومة شريفة
استشاري التحاليل الطبية
Dr. Soma Sherif
Consultant Pathologist

Date

26/3/24