

Amrad



مستشفى العرف
ALORF HOSPITAL

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 12/11/23

Requested by _____

Department Ophthalmology Location IT rel

SR. No.	Machine	Model	Complaint
822897	Nidek RT-3100 Refractor	Nidek RT-3100	Not working
	Digital Refracto.		

Check within Week ASAP

Engineer Comments

Waiting for quotation, need to change the relay
box unit

End Comments

Replaced by new relay box by the company.

Requested by

Checked By

Approved by

مستشفى العرف
1431611064

**Al Amin Medical Company**

Anwar Al-Sabah Complex-Block 2.A-3rd Floor,P.O.Box 23837 Safat-13099 Safat,Kuwait

Tel.:22442034 / 35, Fax: 22403134, Toll Free Number: 22426426

Email: callcenter@amicogroup.com

SERVICE REPORT

Service Order No: 1440026723		Request Start Date: 16.11.2023		Machine Status:	
Customer :	4100249-MANZOOR ALORF CLINIC			Visit Type: REPAIR SERVICE ORDER	
Phone:	97202656	Mobile:	67716055	Country:	Kuwait

Manufacturer:	NIDEK CO.,LTD.	Room No:		Asset No:	
Model:	NIDK RT-3100S REFRACTOR			Serial No:	822897

Problem Description	Date
Check REFRACTOR SN:822897/machine off	

Corrective Action	Date
Checked the device Digital refractor relay box have power issue Need to bereplaced	16.11.2023

Replace/Installed Parts						
Part #	Description'	Qty	Unit Price	Discount	VAT	Total

Date Completed : 16.11.2023	Service Order Status:	Customer Signature 
Time Completed : 16:21:52	Waiting for Customer approval	
Mark Alvin Mortos Engineer	Parts total Amount: 0.00	
		By: Gina