

Roche

QC report

14-10-2024 15:29
 Serial number : 19239
 Instrument ID : 8
 Operator ID :
 ALORF HOSPITAL

QC material : COMBITROL PLUS B

Level : 2

Sample type : Aqueous solution

Lot number : 21431171

Exp. date : 31-07-2025

Equilibration temperature : 0.0 °C

44.2 mmHg	[	40.6 -	47.6]	Ok
104.3 mmHg	[	90.0 -	114.0]	Ok
7.438	[	7.390 -	7.470]	Ok

Roche

QC report

14-10-2024 15:44
 Serial number : 19239
 Instrument ID : 8
 Operator ID :
 ALORF HOSPITAL

QC material : COMBITROL PLUS B

Level : 1

Sample type : Aqueous solution

Lot number : 21431065

Exp. date : 31-05-2025

Equilibration temperature : 0.0 °C

PCO ₂	63.8 mmHg	[	61.2 -	69.2]	Ok
PO ₂	66.5 mmHg	[	50.0 -	74.0]	Ok
pH	7.219	[	7.180 -	7.260]	Ok

Roche

QC report

14-10-2024 15:32
 Serial number : 19239
 Instrument ID : 8
 Operator ID :
 ALORF HOSPITAL

QC material : COMBITROL PLUS B

Level : 3

Sample type : Aqueous solution

Lot number : 21431265

Exp. date : 30-06-2025

Equilibration temperature : 0.0 °C

PCO ₂	27.7 mmHg	[	25.0 -	31.0]	Ok
PO ₂	161.9 mmHg	[	145.0 -	169.0]	Ok
pH	7.547	[	7.520 -	7.580]	Ok

log no-1473

Roche

Date: 14/10/2024		SERVICE REPORT NO: 0212	
Contract No.			
Hospital/Laboratory/Polyclinic AL ORF HOSPITAL		Dept NICU	Case Qualification ☐ Complaint App ☒ Complaint HW ☐ Preventative Maintenance ☐ Installation ☐ Inquiry ☐ Others ☐ Modification ☐ Workshop
Instrument BLOOD GAS ANALYZER		Model Cobas b221	
Serial 19239		SW ver.	
Contact Name		Role	
City	Email	Account ID	
	Tel		
Service Code		Solution	
Contract Type ☐ Warranty ☐ Reagent Rental ☐ Full Service ☐ Maintenance ☐ On call			
Subject TEMPERATURE	Alarm Code	Alarm Description	REXIS Activity # Failure Causing Part
Action: B.G & ISE MC CLEANED. ISE MC DEACTIVATED. INSTRUMENTS TYPE SET TO b221<2> CALIBRATION & GC DONE. ALL 3 LEVELS O.K.			
NOTE HP SINCE MACHINE HAS NO CONTRACT 2 HOURS LABOUR WILL BE CHARGED			
Part #	Description	QTY	Remark CHARGED
	LABOUR HOURS	2	KD. 200/-
INSTRUMENT STATUS			
Job Completed ☒ Yes ☐ No Instrument Working ☒ Yes ☐ No ☐ Under List		DOCUMENTATION Quotation No. Invoice No.	
FIELD SERVICE REP		BIOMED ENGINEER / MOH ENGINEER	
Name: SUNASHY		Name: Ferie Joy Williams	
Signature: [Signature]		Signature: [Signature]	
Date: 14/10/2024		Date: 14/10/2024	
		KRISTINA MARCEUNO ALORF HOSPITAL 4/10/2024	