

MEDICAL EQUIPMENT REPAIR REQUEST

No. #3

Date: 15/01/2025

Requested by Ferre Jry Williams

Department NICU Location 3rd Floor

SR. No.	Machine	Model	Complaint
H0HK51647	Medical Incubator	Giraffe	screen is blurry

Check within

☐ Week

☒

ASAP

Engineer Comments

The screen is working but its blurry and Shady "not clear". Need to Change it with new one.

End Comments

Part #: Display Cage 6600-0222-850
PCALED 6600-0215-850

Ferre

Requested by

Ahmad Alawadh:

Checked By

Carlene Hassan

Approved by



FIELD SERVICE REPORT

LCS 12928

CUSTOMER

Alorf Hospital

JOB NO.

EQUIPMENT

Incubator

MODEL NO.

giraffe ONMEDA

LOCATION

WICU

SERIAL NO.

HDHKS1647

DATE

11 / 3 / 2025

ASSET

CONTRACT

Q12025 - LCS - 104

COST CENTER

4502

BILL PARTS / LABOR ☐

WARRANTY ☐

CONTRACT ☐

INSTALLATION ☐

OTHERS ☒

ENGINEER

HOURS

- Altamimi

- Agmal

- Ahmad

2

PARTS USED

QTY.	PART NUMBER	SIV	DESCRIPTION
1	6600-0222-850		Display cage
1	6600-0215-850		PCA LED

PROBLEM, DIAGNOSIS AND ACTION

* service call attended.

* unit is checked.

* Replace a spare part for the unit.

* unit working in a good condition.

REMARKS

JOB COMPLETED

☒ Yes

☐ No

AWAITING PARTS

☐ Yes

☒ No

ATC ENGINEER SIGNATURE

USER SIGNATURE

Feni Jy Juirianson

AUTHORIZED SIGNATURE & STAMP

KRISTINA MARCELYNO



FIELD SERVICE REPORT

LCS 12906

CUSTOMER ALORF hospital		EQUIPMENT In Cubator		MODEL NO. Giraffe		DATE 19 / 01 / 2025	
JOB NO.		LOCATION NICU		SERIAL NO. BA HDHKS1647		ASSET	
PARTS USED						CONTRACT	
QTY.	PART NUMBER	SIV	DESCRIPTION			COST CENTER 4502	
						BILL PARTS / LABOR ☐	
						WARRANTY ☐	
						CONTRACT ☐	
						INSTALLATION ☐	
						OTHERS ☒	
ENGINEER Ahmed				HOURS 1			

PROBLEM, DIAGNOSIS AND ACTION

- ① Service call attended
- ② Unit checked
- ③ Problem: Unit's screen is Shady and not clear. It is not working well. It has to be change with new screen/display Board
- ④ Repair Request is needed

REMARKS



JOB COMPLETED ☐ Yes ☒ No

AWAITING PARTS ☒ Yes ☐ No

ATC ENGINEER SIGNATURE

USER SIGNATURE

مستشفى العرف
ALORF HOSPITAL

AUTHORIZED SIGNATURE & STAMP

www.atc.com.kw