

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 1773

شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60846

Customer	Hosif Hospital			Equipment	X-Ray		
				Model No.			
				Serial No.			
Authorized Person				Manufacturer	Primax		
Customer Complaint				Location	Radiology		
				Contract / Order No.			
				☐ Installation	☐ Warranty		
				☐ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Services / Check Carried out							
☐ Inspection / Cleaning			☐ Optics Checks				
☐ Calibration / Performance			☐ Pneumatic Checks				
☐ Interface / Communication			☐ Hydraulic Checks				
☐ Safety Checks			☐ Operational Error				
☐ Electrical / Electronics			☐ Mechanical Checks				
☐ Accessories Check			☐				
Remarks	Installed Collimator unit, confirmed, tested and confirmed all functions are working properly.						
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer	Authorized Person			Customer / User			
ATCO Technology Co.							
Signature	Signature			Signature			
Date	Date			Date			