

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60939

Customer	Abraf Hospital.			Equipment	Patient Trolley.		
				Model No.			
				Serial No.	JN50083378-0004		
Authorized Person				Manufacturer	Ummi Medical		
Customer Complaint				Location	4th floor ward.		
head rest not stable.				Contract / Order No.			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
				Time Spent			
Remarks Service call attended. we inspect the unit and found that the head rest is not stable. we adjusted the gas spring for head rest. we tested the head rest in different position. we found that we found that the head rest is stable. we left the unit under observation. Currently the unit is working properly.							
No.	Spare Parts Used		Part No.	Qty.			
Service Engineer		Authorized Person		Customer / User			
Nixon ATCO Technology Co. Signature Date 31/07/2024.		4/8/24 Signature Date		Signature Date			