

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____			
Date of Occurrence: <u>29/05/25</u>	Time of Occurrence: <u>10am</u>	Occurrence Location: <u>4th Floor (4B)</u>	
If event involved employee - Staff Name: <u>Teenand Ugeel, Neethu M L</u>		Age: _____	Sex: _____ Emp#: _____
If event involved patient - Patient Name: _____		Age: _____	Sex: _____ File#: _____
If event involved Visitor - Visitor Name: _____		Phone No.: _____	
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☐ Not Working ☐ Improper Use ☒ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>Battery failure</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE		
☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: _____	ID No. _____	Dept. _____	Position _____
IMMEDIATE ACTION				
Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				

Assessment/ Diagnosis			
Physician Name:	Dept.:	ID No.:	

DUPE

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

INCIDENT REPORT FORM

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

When we checked defibrillator on 04/06/25 machine showed error in the monitor and we are informed biomedical staff and she said that it was battery failure

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
☐ **Moderate.** Increased length of Stay. Increased level of care.
☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
7 referred to Biomed		- Referred to Biomedical for immediate inspection.	
Dept. Head: Christina Mena	Date: 4/6/25	Quality Mgmt. Dept. 	Date: 
ID No. 1022	Department: Nursing	Comments:	

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 04.06.25

Request by Teenamol Gebeil, Neethu M C

Department Dermatology Location 4th Floor (4B)

SR. No.	Machine	Model	Complaint
<u>212060645</u>	<u>Defibrillator</u>	<u>Cardio-Arr 360-B STARTTM Biphasic</u>	<u>Battery failure</u>

Check within Week

☒ ☐ ASAP

Engineer Comments

End Comments

Request by

Checked by

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 59800

Customer	Alort Hospital			Equipment	ENT Defibrillator		
				Model No.	Cardia aid 360 B		
				Serial No.	212060645		
Authorized Person				Manufacturer	Innowell		
Customer Complaint				Location	4th Floor		
Battery not working				Contract / Order No.			
Services / Check Carried out				☐ Installation ☐ Warranty			
☐ Inspection / Cleaning				☐ Repair			
☐ Optics Checks				☐ On Charge			
☐ Calibration / Performance				☐ Maintenance			
☐ Pneumatic Checks				☐ Contract			
☐ Interface / Communication				☐ Upgrade			
☐ Hydraulic Checks				☐ _____			
☐ Safety Checks				Time Spent			
☐ Operational Error							
☐ Electrical / Electronics							
☐ Mechanical Checks							
☐ Accessories Check							
☐ _____							
Remarks Done Battery format and reset battery setup. Unit under observation for charging. It still not working after battery format, Battery need to be replaced.							
No.	Spare Parts Used			Part No.	Qty.		
Service Engineer				Authorized Person		Customer / User	
Signature <i>[Signature]</i> Date 9/6/25				Signature <i>[Signature]</i> Date 10-06-2025		Signature _____ Date _____	