

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



ATCOTECH

Log no - 1783

شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 53624

Customer	Al-Ory Hospital Kuwait	Equipment	Digital X-ray system
		Model No.	NZLCA - C7575
		Serial No.	
Authorized Person		Manufacturer	PRZMAX INTERNATIONAL
Customer Complaint	Touch Screen Faulty	Location	X-ray Dept
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
		Time Spent	

Remarks

Replaced Touch Screen CPU Board.
Checked and Recalibrated System with
online support from Premax.

Checked all operations found OK

Note: Unit is old, software and hardware
support limited and future support

No.	Spare Parts Used	Part No.	Qty.
	can be guaranteed		

Service Engineer

Azeem + Syed

ATCO Technology Co.

Signature

Date

24/7/24

Authorized Person

Signature

Date

24/7/24

Customer / User

Signature

Date

24/7/24