

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no- 1775

شركة اتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58551

Customer	Al-Ory Hospital Kuwait	Equipment	Digital Mammography System
		Model No.	Helianthus - C
		Serial No.	
Authorized Person		Manufacturer	Metal Trexco
Customer Complaint	Compression paddle not released after exposure.	Location	Radiology Dept
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐		
		Time Spent	

Remarks
Checked system, re-enabled compression release activation
checked all operations found ok

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Amm. K ATCO Technology Co. Signature Date 27-2-24	Authorized Person Signature Date 2/3	Customer / User Signature Date 2/3
--	--	--