

P.O.Box : 24240 Safat
Code : 13103 | Kuwait
C.R Number : 353520
Capital 500,000 K.D



ص.ب : 24240 الصفاة
الرمز : 13103 | الكويت
السجل التجاري : 353520
رأس المال 500,000 د.ك

SERVICE REPORT

NO. 1058

SERVICE CALL NO.

Customer's name & address <u>Al Ory Hospital</u>	Equipment <u>CT system / CT injector</u>
Contact person	Serial No. <u>59412 / 860 013 825</u>
Department <u>Radiology</u>	Model No. <u>Somatom Perspective / Accutron CT-D</u>
Tel. No.	Manufacturer <u>Siemens / Medtron</u>
	Maint. Cont. _____ Ord. No. _____
	Inv. No. _____ Date <u>10-02-2025</u>

INSTALLATION ☐ WARRANTY ☐ PREVENTIVE MAINTENANCE ☐ CHARGEABLE ☐ REPAIR ☒ UPGRADE ☐ UPDATE ☐

Description of WORK DONE :

* Nacl Piston side was not functioning after changing the Nacl Motor it solved the problem
* New Housing Cover also replaced to prevent from liquid dropping

Repair Completed Yes ☒ No ☐

REMARKS

Eng	Date	Time		HRS.		Over Hrs.
		From	To	Work	Wait	Trav.
OJ ABD	10-02-2025					

SPARE PARTS USED:

Description	Article No.	Stock No.	Qty.
1- CT-D Motor (Nacl)	E 60 1178		1
with Bracket, coupling and Rotary encoder disc (2)- Driving Mechanism Cover CT-D	E 606 101 B		1

Parts to be Ordered :	Article No.	Description:	Qty:

Service Engineer's name & Signature

Date 10-02-2025

CUSTOMER's name & signature

Date

APPROVED BY:

Date