

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



log no -

شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجاهراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63215

Customer	Morif Hospital	Equipment	Patient Monitor.
		Model No.	Em3
		Serial No.	26143 - M19C0293-
Authorized Person		Manufacturer	Edan ٥٥٥١
Customer Complaint		Location	NICU.
		Contract / Order No.	
	BP Abnormal.		
Services / Check Carried out			
☐ Inspection / Cleaning	☐ Optics Checks	☐ Installation	☐ Warranty
☐ Calibration / Performance	☐ Pneumatic Checks	☒ Repair	☐ On Charge
☐ Interface / Communication	☐ Hydraulic Checks	☐ Maintenance	☐ Contract
☐ Safety Checks	☐ Operational Error	☐ Upgrade	☐ _____
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Time Spent		
suspected unit interchanged bp cuffs, confirmed unit & wiring properly.			

No.	Spare Parts Used	Part No.	Qty.

Service Engineer ATCO Technology Co. Signature Date 3/11/24	Authorized Person Signature Date	Customer / User Fare Joy Williams Signature Date
--	--	---