

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة اتكو التكنولوجيه

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 59275

Customer	Alorad Hospital Jahra		Equipment	Patient Monitor	
			Model No.	VITADIA 7000R	
			Serial No.	05BM000271	
Authorized Person			Manufacturer	Trismed	
Customer Complaint			Location	E.P	
Not charging			Contract / Order No.		
Services / Check Carried out			Time Spent		
☒ Inspection / Cleaning			☐ Optics Checks		
☒ Calibration / Performance			☐ Pneumatic Checks		
☒ Interface / Communication			☐ Hydraulic Checks		
☐ Safety Checks			☐ Operational Error		
☐ Electrical / Electronics			☐ Mechanical Checks		
☐ Accessories Check			☐		
Remarks			Service call attended. Checked and found the battery need to reset. Reset the battery module. Checked and found it working properly.		
No.	Spare Parts Used		Part No.	Qty.	
Service Engineer			Authorized Person		Customer / User
Signature			Signature		Signature
Date			Date		Date