

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58342

Customer	Alor Hospital		
Equipment	Autoclave		
Model No.	STERILCLAVE 18		
Serial No.	13X 22079		
Authorized Person	Manufacturer Cominox		
Customer Complaint	Location 1st floor,		
Steam leakage.	Contract / Order No.		
Services / Check Carried out	☐ Installation ☐ Warranty		
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐		
Remarks	Service call attended. While doing the cycle the steam is leak from the bottom part of door. We adjusted the door alignment and tested. We successfully run a cycle and we couldn't find any leaks. The unit is working properly.		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Miscom			
ATCO Technology Co.			
Signature	Signature	Signature	
Date 29/02/2024	Date	Date	