

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب. ٢٣٥٥ الجهرء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63843

Customer	Alor Hospital	Equipment	Autoclave
		Model No.	selectomat PL
		Serial No.	220823
Authorized Person		Manufacturer	MMN
Customer Complaint		Location	CSSD
		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☒ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐	
	Time Spent	

Remarks	Service call attended he inspected the unit and tested one cycle he couldn't find any stain on the packing. he found that the before cycle which cause stain having more than 5hr time gap. so we recommend the user to do pre heat cycle before using the machine after long gap.
---------	--

No.	Spare Parts Used	Part No.	Qty.
	currently the unit is working properly.		

Service Engineer M. Al-Sayid ATCO Technology Co. Signature Date 18/03/2025	Authorized Person Signature Date 27/3/25	Customer / User Signature Date
--	--	--------------------------------------