

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2167

شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60899

Customer Al-Fal Hospital Jahra	Equipment ESWL unit
	Model No. Integra
	Serial No. 118
Authorized Person	Manufacturer Omrex
Customer Complaint	Location OR
	Contract / Order No.

Services / Check Carried out	☐ Installation	☐ Warranty
	☒ Repair	☒ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐ _____
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐ _____	
	Time Spent	

Remarks
Charged the water with new Removed air bubbles from the bellows. Performed maintenance for the unit and done test. Checked and found that the unit is working properly

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Sunth ATCO Technology Co. Signature Date 26/08/2024	Authorized Person [Signature] Signature Date 27/8/24	Customer / User Signature Date
---	---	--