

Customer Aloof Hospital		Equipment Autoclave	
		Model No. Selectomat PL	
		Serial No. 220824	
Authorized Person		Manufacturer mmn	
Customer Complaint Stain on instrument case		Location CSSD	
		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Optics Checks ☐ Calibration / Performance ☐ Pneumatic Checks ☐ Interface / Communication ☐ Hydraulic Checks ☐ Safety Checks ☐ Operational Error ☐ Electrical / Electronics ☐ Mechanical Checks ☐ Accessories Check ☐		☐ Installation ☐ Warranty	
		☒ Repair ☐ On Charge	
		☐ Maintenance ☐ Contract	
		☐ Upgrade ☐	
		Time Spent	
Remarks Service call attended, we inspect the unit and found that the drainage is blocking with the impurities in steam generator. we cleaned the Steam generator and forced it properly. Checked and confirmed that the unit is working properly.			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer Nishan/mouse ATCO Technology Co. Signature (f) Date 24/08/2024.		Authorized Person 26/08/24 Signature Date	
Customer / User Signature Date			