

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____	
Date of Occurrence: <u>4/3/25</u> Time of Occurrence: _____ Occurrence Location: <u>5B</u>	
If event involved employee - Staff Name: <u>SHIBY C. PHILIP</u> Age: <u>37</u> Sex: <u>F</u> Emp#: <u>4155</u>	
If event involved patient - Patient Name: _____ Age: _____ Sex: _____ File#: _____	
If event involved Visitor - Visitor Name: _____ Phone No.: _____	
Address: _____	

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☐ Not Working ☐ Improper Use ☒ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other _____

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE		
☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: <u>SHIBY C. PHILIP</u>	ID No. <u>4155</u>	Dept. <u>5B</u>	Position <u>Staff</u>
IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				

Assessment/ Diagnosis		
Physician Name:	Dept.:	ID No.:

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

We Found Bed siderails Broken on 4/7/25 11 PM Room-518 after discharge of patient. Room 513 Bed also Complaint head part is not coming up.

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
Referred to Biomed		Referred to Biomed for repair or replacement.	
Dept. Head:	Christina Chan	Quality Mgmt. Dept.	Stavros TB RN
Date:	6/7/25	Date:	
ID No.	1022	Department:	Nursing
		Comments:	

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 5/7/25

Requested by Shiby C Philip

Department ward Location 5B

SR. No.	Machine	Model	Complaint
<u>380-2008196</u>	<u>Bed</u>		<u>Headend is not working sound also coming.</u>

Check within Week

ASAP

Engineer Comments

End Comments

Requested by _____

Checked By _____

Approved by _____

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 64667

Customer	ALORF HOSPITAL	Equipment	Medication bed
	Jahra	Model No.	S380 Sounder
		Serial No.	3802008-44-196
Authorized Person		Manufacturer	Volker
Customer Complaint	Not working	Location	Room No: 543
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☒ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ EM
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐	Time Spent	
Remarks	Checked found control box and micro switch defective. Replaced control switch and micro switch All connectors are tested OK Bed working good. ① Technical issue internal to the device		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Tarek Al-Chorani			
ATCO Technology Co.			
Signature	Signature	Signature	
Date	Date	Date	
06/07/2025	6/7/2025		