

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____			
Date of Occurrence: <u>6/5/25</u>	Time of Occurrence: <u>3:00 PM</u>	Occurrence Location: <u>LAB 142A</u>	
If event involved employee - Staff Name: <u>ALFREDA SABBAGH</u>	Age: <u>50</u> Sex: <u>F</u> Emp#: <u>142A</u>		
If event involved patient - Patient Name: <u>(LABORATORY)</u>	Age: _____ Sex: _____ File#: _____		
If event involved Visitor - Visitor Name: _____	Phone No.: _____		
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR ☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	SECURITY VARIANCE ☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	EQUIPMENT/SUPPLIES ☐ Electrical Problem ☒ Not Working ☐ Improper Use ☐ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>Centrifuge</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☒ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: <u>Dr. Anyth</u>	ID No.	Dept.	Position
IMMEDIATE ACTION				
Physician Notified	☐ Yes ☐ No	Did MD see patient	☐ Yes ☐ No	☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication

Assessment/ Diagnosis			
Physician Name:	Dept.:	ID No.:	

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

The Centrifuge still working @ 3pm today (6/5/24) but @ 3:05pm, when we Centrifuge 2 samples, the Centrifuge Suddenly Stop. The motor stop and stop functioning, as well!

Stop: 3pm (6/5/24)

NOT Functioning: 3:05pm (6/5/24)

SN: 001900026

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
Inform Biomedical department		Referred to Biomedical for repair.	
Dept. Head: Dr. Anwar Sami		Quality Mgmt. Dept. [Signature]	
Date: 12-5-2025		Date:	
ID No.	Department:	Comments:	

Department Managers: Forward Completed forms to Quality Management Department



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 6-5-2025

Requested by Lab. department . Dr. Arafat Sami

Department lab Location _____

SR. No.	Machine	Model	Complaint
<u>601900026</u>	<u>Centrifuge</u>	<u>Hermle</u>	<u>not working</u>
		<u>Z 206 A</u>	

Check within

Week

☒

ASAP

Engineer Comments

End Comments

Dr. Arafat Sami

Requested by

6.5-2025

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63010

Customer	ALORF Hospital	Equipment	Centrifuge
	Lab Jahra	Model No.	Z 206 A
		Serial No.	601900026
Authorized Person		Manufacturer	Heamle
Customer Complaint	Not working	Location	Lab
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ EM
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐		
		Time Spent	

Remarks

Checked found Centrifuge lid switch defective.
Need to replace the switch, informed customer
Waiting for the spare parts

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

Inice Cherian
ATCO Technology Co.

Signature

Date 08/05/2025

Authorized Person

Signature

Date

Customer / User

Dr. Anfal Samy
8-5-2025

Signature

Date

22023120560

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63012

Customer	ALORF Hospital	Equipment	Centrifuge
	Jahra	Model No.	Z 206 A
		Serial No.	601900343
Authorized Person		Manufacturer	Hermle
Customer Complaint	not working	Location	Laboratory dept.
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ <u>EM</u>
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		

Remarks

Checked fanned lid switch defective.
Informed customer need to replace the switch
Waiting for the spare parts to solve this problem

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

Tince Cherian
ATCO Technology Co.

Signature

Date 08/05/2025

Authorized Person

Signature

Date

Customer / User

Dr. Amlat Sami
8-5-2025

Signature

Date ٨ ايات سامي محمد عبدالله
طبيب مستشفى الجهرام
مستشفى الجهرام
22023120560