

Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218



تلفون: ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)
فاكس: ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)
ص.ب: ٢٢٥٥ الجھراء، ٠١٠٢٥ - الكویت
Email: sales@atco-kuwait.com
www.atco-kuwait.com
س.ت. ٣٣٦٢١٨

No.: 55712

SERVICE / INSTALLATION REPORT			
Customer Alor Hospital	Equipment Endoscopy washer		
	Model No. SERIE -3		
	Serial No. 25186		
Authorized Person	Manufacturer Soluscope		
Customer Complaint	Location Endoscopy		
	Contract / Order No.		
☐ Installation ☒ Repair ☐ Maintenance ☐ Upgrade ☐ Warranty ☐ On Charge ☐ Contract ☐ _____			
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐ _____			
Time Spent			
Remarks Service call attended.			
We inspect the unit and found error 12-30 while run perform cycle. we changed BNRV for the solution injection line and changed all connectors to the solution can. Then we checked the solution injection pump. we found that the pump is faulty. need to replace the pump.			
No.	Spare Parts Used	Part No.	Qty.
	Currently the unit is not working.		
Service Engineer Midan ATCO Technology Co. Signature [Signature] Date 27/07/2024		Authorized Person Signature _____ Date _____	
Customer / User Signature _____ Date _____			