

INCIDENT REPORT FORM

LOG #:

Forward Completed Form to Department Manager. One copy to Quality Management Department.

PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☐ Other (specify) _____

Date of Occurrence: 23/6/2024 Time of Occurrence: _____ Occurrence Location: CSC

If event involved employee - Staff Name: Remya Age: _____ Sex: _____ Emp#: 4231

If event involved patient - Patient Name: _____ Age: _____ Sex: _____ File#: _____

If event involved Visitor - Visitor Name: _____ Phone No.: _____

Address: _____

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR	SECURITY VARIANCE	EQUIPMENT/SUPPLIES
☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	☐ Electrical Problem ☒ Not Working ☐ Improper Use ☐ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other <u>CSSD WASHING POOL</u>

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE

☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)		Response to Treatment/Outcome:
☐ Wrong route		
☐ Wrong dose		

Reported by Name: _____ ID No. _____ Dept. _____ Position _____

IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication

Assessment/
Diagnosis

Physician Name:

Dept.:

ID No.:

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

- 1 Today we start Washer II But after the washing door not open.
- 2 Autoclave II have white spots Top of the Instruments after Autoclaved.

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT	REFERRED TO QUALITY MANAGEMENT DEPT.
3 referred to HMTS	
4 per by the company HMTS, he discussed why the machine is always has problem with the door being stuck up for Feb, March & May. for the #1 WASHER. they are changing the monthly wheels & the door is not fair	
5 WASHER #2 always automatically not working	
Dept. Head: <u>Christina Guevara</u> Date: <u>24/6/2015</u>	Quality Mgmt. Dept. Date:
ID No. <u>1022</u> Department: <u>Nursing</u>	Comments:

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 23/6/25

Requested by Remya

Department CSSD Location CSSD.

SR. No.	Machine	Model	Complaint
<u>230506</u>	<u>washer disinfecter</u>	<u>MMM</u>	<u>door not open</u>

Check within Week

ASAP

Engineer Comments

End Comments

Remya
Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63753

Customer	ALORF HOSPITAL	Equipment	Washer & Disinfectant
	Jahra	Model No.	Uniclear
		Serial No.	230506 (#2)
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	CSSD
Unloading door not opened		Contract / Order No.	

Services / Check Carried out

- | | |
|--|--|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ |

- | | |
|---------------------------------------|------------------------------------|
| ☐ Installation | ☐ Warranty |
| ☐ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ <u>EM</u> |

Time Spent

Remarks

Found Unloading door not opened.
Reset door delay and reset the program.
One cycle successfully completed. All connectors are tested OK. System working good.
① Door ~~was~~ not opened due to ~~delay~~ technical issue system under observation.

No.	Spare Parts Used	Part No.	Qty.
	ON 9/6/25 checked AND		
	RESET done, working		
	after RESET.		

Service Engineer Jince Cherian ATCO Technology Co. Signature <u>[Signature]</u> Date 23/06/25	Authorized Person <u>[Signature]</u> 28/6/25 Signature Date	Customer / User Signature Date
---	---	--------------------------------------