

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكوتكنولوجيه

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63785

Customer	Alorif Hospital			Equipment	OT Table		
				Model No.	V750		
				Serial No.	NV111620091		
Authorized Person				Manufacturer	NUVO		
Customer Complaint	oil leakage			Location	OT Room 5		
				Contract / Order No.			
Services / Check Carried out	☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check			☐ Installation ☒ Repair ☐ Maintenance ☐ Upgrade	☐ Warranty ☐ On Charge ☐ Contract ☐		
	☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐			Time Spent			
Remarks	Service call attended. we inspect the unit and found that the oil is leaking. we cleaned the base and found that there is one tube & cap is loose. we tighten the cap and fixed it properly. checked and confirmed that the unit is working properly. we need to fill the unit with hydraulic oil. next time we will refill.						
No.	Spare Parts Used		Part No.	Qty.			
	Oil. next time we will refill.						
	NOT working properly need replacement of oil						
Service Engineer			Authorized Person		Customer / User		
Niscom ATCO Technology Co.							
Signature			Signature		Signature		
Date 22/02/2024			Date 2/11/25		Date		