

INCIDENT REPORT FORM

LOG #:	Forward Completed Form to Department Manager. One copy to Quality Management Department.		
PART A. Status: ☐ Inpatient ☐ Outpatient ☐ Visitor/Family ☐ Employee ☒ Other (specify) <u>Machine</u>			
Date of Occurrence: <u>07/10/2015</u>		Time of Occurrence: <u>0900H</u> Occurrence Location: <u>NW</u>	
If event involved employee - Staff Name: <u>Ferre Jny Williams</u>		Age: _____ Sex: _____ Emp#: <u>1465</u>	
If event involved patient - Patient Name: _____		Age: _____ Sex: _____ File#: _____	
If event involved Visitor - Visitor Name: _____		Phone No.: _____	
Address: _____			

PART B. Occurrence Type (check boxes that apply)

INTRAVENOUS RELATED	PROCEDURE VARIANCE	HAZARD/SAFETY
☐ Central Line ☐ Contrast Media Reaction ☐ Extravasate ☐ Occluded ☐ Precipitate ☐ Site Red/Swollen ☐ Tubing Leak ☐ Expired IV Fluid ☐ Additive Related ☐ Transfusion Reaction ☐ Other _____	☐ Delays ☐ Missing Files, Notes ☐ Performed on Wrong Site ☐ Consent ☐ Patient Identification ☐ Wrong Patient ☐ Wrong Procedure ☐ Lost/Spoiled Specimen ☐ No Response ☐ Wrong File ☐ Procedure Cancellation ☐ Notes in Wrong File ☐ Inadequate/Inappropriate Entry ☐ No label/Addressograph ☐ Other _____	☐ Burn ☐ Chemical Spill ☐ Electrical Shock ☐ Fire ☐ Gas ☐ Hazardous Waste ☐ Radiation Exposure ☐ Splash ☐ Water Spill ☐ Needle prick injury ☐ Other _____
BEHAVIOR ☐ Improper disposal of sharp ☐ AMA/Absconded ☐ Dissatisfied with care ☐ Patient Behavior ☐ Family/Visitor Behavior ☐ Staff Behavior ☐ Time In/Time Out Variation ☐ Non-Compliance ☐ Patient Complaint ☐ Other _____	SECURITY VARIANCE ☐ Damage/Loss of Property ☐ Kidnapping ☐ Missing/Misplaced Drug Keys ☐ Prescription Alteration ☐ Security Problem ☐ Suicide attempt ☐ Other _____	EQUIPMENT/SUPPLIES ☒ Electrical Problem ☐ Not Working ☐ Improper Use ☐ Mechanical Problem ☐ Operator Not Qualified ☐ Wrong Equipment ☐ Missing Equipment ☐ Not Available ☐ Out of Stock ☐ Other _____

ADVERSE DRUG REACTION (ADR) / MEDICATION OCCURENCE		
☐ Medication missing	☐ Wrong Patient	Describe Therapy:
☐ Adverse Side effects	☐ Others:	
☐ Medication given but not charted		
☐ Medication charted but not given	*please see Intravenous Related area	
☐ Duplication/extra dose given	Describe reaction (chronological sequence):	
☐ Time variance (wrong time)	Response to Treatment/Outcome:	
☐ Wrong route		
☐ Wrong dose		

Reported by	Name: <u>Ferre Jny Williams</u>	ID No. <u>1465</u>	Dept. <u>NW</u>	Position <u>Head Nurse</u>
IMMEDIATE ACTION Physician Notified ☐ Yes ☐ No. Did MD see patient ☐ Yes ☐ No ☐ X-ray ☐ Suture ☐ Blood Work ☐ Medication				

Assessment/ Diagnosis			
Physician Name:		Dept.:	ID No.:

INCIDENT REPORT FORM

"An Incident" is any happening that is not consistent with the routine operation of the hospital or the routine care of a particular patient which adversely affects or threatens the health or life of a patient, visitor or employee. It may be an occurrence or incident which might result in an accident.

PART C. Describe the occurrence (state the facts). Use when necessary.

FACTUAL DESCRIPTION

While doing routine checking, noticed the transport incubator was shut being charged. When unplugged, there was an alarm that there is no battery / power. tried to plugged in another port / area and after an hour not charging. Kaze & Engr. Benil was informed

PART D. Severity Codes (Please check one).

- ☐ **Minor.** No injury or minor treatment, i.e. abrasion, No increased level of care.
- ☐ **Moderate.** Increased length of Stay. Increased level of care.
- ☐ **Major.** Temporary/permanent lessening of bodily functions (sensory, motor, physiologic, or intellectual). Not related to the natural Outcome of the patient's illness, underlying condition, or disfigurement. Surgical intervention required.
- ☐ **Sentinel/Catastrophic.** Death/Major permanent loss of function (sensory, motor, physiologic, or intellectual), hemolytic transfusion reaction, surgery on wrong patient or body part.
- ☐ **Near Miss.** For potentially severe Outcome.

PART E. ACTIONS.

PREVENTIVE ACTION

ACTION TAKEN BY INVOLVED DEPARTMENT		REFERRED TO QUALITY MANAGEMENT DEPT.	
Reported to Biochem			
7 finger ends to come			
Dept. Head: <u>Christina A. A. A.</u> Date: <u>7/7/2025</u>		Quality Mgmt. Dept. Date:	
ID No. <u>1022</u> Department: <u>Nursing</u>		Comments:	

Department Managers: Forward Completed forms to Quality Management Department

MEDICAL EQUIPMENT REPAIR

No. 3

Date: 09/07/2025

Requested by Fayre Jy miranem

Department NICU Location 3rd Floor

SR. No.	Machine	Model	Complaint
<u>2261-16</u>	<u>Transport Incubator</u>	<u>TR 200</u>	<u>Battery not charging</u>

Check within

Week

☒ ASAP

Engineer Comments

End Comments

Fayre

Requested by

Checked By

Approved by

Sl. No: 42024

Call No:		Call Received Date:		Call Received Time:	
Customer Name & Address: Albra Hospital Kunbar		Date and Time Reported to client Place:			
		Engineer Name		Date	
		Bini George Bini Skara		08/07/25	
Contact Person Name:					
Email ID:			Mobile #:		
Contract Type:	☐ Warranty	☐ LAMC	☐ CAMC	☐ HBS	☐ Others
Call Type:	☐ Installation	☐ PMS	☐ Breakdown	☐ Site Visit	☐ Inspection
Call Completed Date & Time:			Contract / Order No.		
Equipment Manf: Medix			Eqpt. Model: TR-300		
Eqpt. Type: Transport Incubator					
Sl.No. 2851-16	Location:		S/W. Version:		
Customer Complaint:					
Service Performed: Checked the unit and found that the Battery is faulty. Need to change the faulty Battery.					
Slice Count:		Rotation Count:		mAs Count:	
Suggestions to Customer:				Customer's Signature & Seal	
SPARE REPLACED					
Sl. #:	Part Name / Description:	Part Code / Part No. :	Qty.		
Customer's Comments:				For my insurance	
Engr. Name: Bini George		Engr. Signature: Bini George		Date: 08/07/25	

Gulf Advanced Trading Co. W.L.L

شركة تقدم الخليج التجارية ذ.م.م

Al Awadi Tower, Tower-1, Floor No. 2, Ahmed Al-Jaber Street, Sharq, Kuwait

Tel : (965) 22473437 Fax: (965) 22473436 Service@gatkuwait.com www.gatkuwait.com