

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2591

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61978

Customer	Alorif Hospital	Equipment	Autoclave
		Model No.	Selectomab PL
		Serial No.	220823
Authorized Person		Manufacturer	mmm
Customer Complaint		Location	CSSD
	B&w test failure	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		

Remarks

Service call attended.
we inspect the unit and found that the unit is
not achieving proper vacuum.
we adjusted the anti cavitation valve
now the unit is working properly. checked and
confirmed that the unit is working properly.

No.	Spare Parts Used	Part No.	Qty.

under warranty

Service Engineer	Authorized Person	Customer / User
vision/swith/mouse		
ATCO Technology Co.		
Signature	Signature	Signature
Date 17/08/2024	Date	Date