

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (+٩٦٥)

فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (+٩٦٥)

ص.ب : ٢٣٥٥ الجبراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60929

Customer	Aloof Hspat.4.			Equipment	ECG machine		
				Model No.	SE 1200 Express		
				Serial No.	M16606060001		
Authorized Person				Manufacturer	Edan		
Customer Complaint				Location	floor 6		
				Contract / Order No.			
				☐ Installation	☐ Warranty		
				☒ Repair	☐ On Charge		
				☐ Maintenance	☐ Contract		
				☐ Upgrade	☐		
Services / Check Carried out				Time Spent			
☐ Inspection / Cleaning	☐ Optics Checks						
☐ Calibration / Performance	☐ Pneumatic Checks						
☐ Interface / Communication	☐ Hydraulic Checks						
☐ Safety Checks	☐ Operational Error						
☐ Electrical / Electronics	☐ Mechanical Checks						
☐ Accessories Check	☐						
Remarks	we inspect the unit and tested the ECG using Stimulator. we found that the cable is faulty. we replaced the faulty cable with new one in stock. checked and confirmed that the unit is working in good condition						
No.	Spare Parts Used		Part No.	Qty.			
Service Engineer	Authorized Person		Customer / User				
Nison	14/7/2024						
ATCO Technology Co.							
Signature	Signature		Signature				
Date	Date		Date				
08/07/2024							