

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 1775 شركة اتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهره , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58575

Customer	Aloft Hospital Jahra	Equipment	Mamogram
		Model No.	
		Serial No.	
Authorized Person		Manufacturer	Metalsonica
Customer Complaint	Net working	Location	X-Ray Dept
		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☒ Repair	☒ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐ _____
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐ _____	

Remarks	Service call attempted. Checked the unit and found the keyboard and mouse for the PC system not communicating. Reset the power for the USB receiver HUB. Checked and found that the unit is working properly.
---------	---

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Signature: <i>Sujith</i> ATCO Technology Co. Signature: <i>[Signature]</i> Date: 12/02/2024	Authorized Person Signature: <i>[Signature]</i> Date:	Customer / User Signature: <i>[Signature]</i> Date: 12/02/2024
--	---	--