



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 13/01/2025

Requested by Shakendra Yadav

Department PT-2, Male Location Ground floor, Male PT Dept

SR. No.	Machine	Model	Complaint
<u>T6313C</u>	<u>Hydrocollator</u>		<u>The power goes out most of the time I turn the hydrocollator on/off</u>
<u>L20030006-5019</u>	<u>Voltage Regulator</u>		<u>" "</u>

Check within Week

ASAP

Engineer Comments

End Comments

Shakendra Yadav

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء , ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 64307

Customer	Aloraj Hospital	Equipment	Hydro collator
	Physiotherapy	Model No.	2302
		Serial No.	T 6313C
Authorized Person		Manufacturer	
Customer Complaint	Not working	Location	P.T DEPT.
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☒ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☒ EM
☒ Safety Checks	☐ Operational Error	Time Spent	
☒ Electrical / Electronics	☐ Mechanical Checks		
☒ Accessories Check	☐		
Remarks	Power goes off		
	Replace voltage regulator and power plug		
	All connections are checked and system working great		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Jinca Chenian			
ATCO Technology Co.			
Signature	Signature	Signature	
Date 14/01/2025	Date	Date	