

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61989

Customer	Aloft Hospital.	Equipment	Benchtop Autoclave
		Model No.	Storidare 24
		Serial No.	18 x 25721
Authorized Person		Manufacturer	cominox
Customer Complaint		Location	Dental.
Error chamber level		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☒ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Maintenance	☐ Contract
☐ Calibration / Performance	☐ Upgrade	☐ _____
☐ Interface / Communication	Time Spent	
☐ Safety Checks		
☐ Electrical / Electronics		
☐ Accessories Check		
☐ Optics Checks		
☐ Pneumatic Checks		
☐ Hydraulic Checks		
☐ Operational Error		
☐ Mechanical Checks		
☐ _____		

Remarks	Service call attended. we inspect the unit and found that the chamber level error is cleared the level sensor and the filter. we tested one cycle and successfully done. checked and confirmed that the unit is working properly
---------	--

No.	Spare Parts Used	Part No.	Qty.

Service Engineer ATCO Technology Co. Signature Date 27/05/2024	Authorized Person Signature Date	Customer / User Signature Date طبيب أسنان مسجل أول الاستقامة المهنية مستشفى السيف 22019100056
---	--	--