

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



log no-

شركة اتكو التكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 57825

Customer	Aloof Hospital	Equipment	Vital Sign Monitor
		Model No.	Vifapia 5000V
		Serial No.	OSTAF 000 643
Authorized Person		Manufacturer	Trasmed
Customer Complaint		Location	
	NIBP come back broken	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Repaired broken NIBP Connector. Unit is working properly.		
Note : Pls. check / change NIBP hose / cuff.			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
ATCO Technology Co.			
Signature	Signature	Signature	
Date	Date	Date	
17/1/24			