

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2591

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60889

Customer	Alorof Hospital	Equipment	Steam Sterilizer
		Model No.	Selebomat
		Serial No.	220823
Authorized Person		Manufacturer	MMM
Customer Complaint	Clean side door stuck	Location	CSD
		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☐ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐	
	Time Spent	

Remarks	Service call attended. Checked the unit and found the trolley was not inserted properly and this was due to mis handling. Please instruct the user to take care and handle in proper way.
---------	---

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Signature Date	Authorized Person Signature Date	Customer / User Signature Date
Singh ATCO Technology Co. 18/09/2024	 19/9/24	