

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 24/04/24

Requested by Amulika K

Department LDR Location _____

SR. No.	Machine	Model	Complaint
<u>19770208</u>	<u>Infusion pump</u>	<u>Fresenius</u>	<u>Alarming</u>

Check within

☐ Week

☐ ASAP

Engineer Comments

after the evaluation found unit showing continuous
alarm, unit taking to company for repair

End Comments

Repaired now battery

Amulika K

Requested by

Teena

Checked By

Approved by

SERVICE REPORT

-- 1390

CUSTOMER NAME & ADDRESS Al-ORF Hospital		Date: 29/4/2024	
		Equipment: Infusion pump	
		Manufacture: Fresenius	
		Model No: vial optima PT1S	
		Serial No: 19719813, 19770208	
FAULT REPORT / SERVICE DETAILS: Not working			
SERVICE CLARIFICATION: ☐ INSTALLATION ☐ WARRANTY ☐ PMS ☒ OTHERS ☐ ANNUAL MAINTENANCE CONTRACT			
Work Hour	From	To	Total Working Hours
Work Date			Total Working Days
SERVICE PERFORMED			
After the evaluation found the unit's showing continuous alarm. need to take the unit to our office for further evaluation and repair.			
☐ Completed ☒ Under Observation			
CUSTOMER REMARKS:			
Job Satisfactorily Completed ☐ YES ☐ NO		For Cyrix & TSL Company	
Customer Signature: 		Engineer Name: Jawahar A Signature: 	