

Roche for the Trade in Medicines, Equipment, Devices
and Medical Supplies SPC
Customer Report Form



Case number CAS-0018613655

Order number WO-20282509 | Visit Date 10/09/2023

Instrument Details

Instrument/Module cobas b 221<6>=Roche OMNI
S6 system
Serial Number 19239

Customer Details

Customer Number 0050250925
Customer Name Dr. Mansour Alorf Hospital
Street Address Mansour Saad Al Bannaq
Street
Zip 01025 - Jahra
Contact Name Gina Bisenio
Contact Phone +96565991613

Additional Details

PO

Service Activity Code Service Visit

Purpose of Visit
SD rinse and Bypass Rinse errors

Performed Activities
Rinse bottle empty.
New rinse bottle installed.
Rinse okay.
Calibration and QC done and passed.

Time Report


Category	Start	End	Hours
05913616001-Service labour time	10/09/2023 11:00 am	10/09/2023 12:00 pm	1.00

Signature or Stamp

Customer
Gina Bisenio

Roche Support
Sunash Shukoor

Sign Here



By signing this document

- Roche Support person confirms the above mentioned work and checks have been completed..
- Customer representative confirms they have reviewed the work carried out as described and accepts the system for use.
- Thank you for your trust.

Address: Al Mirqab, Al Soor Street, Block 1, Burj Al Shaya, Plot 6, Floor 5, Office
11. P.O. Box: 467, Safat - 13005 Kuwait
Hotline: 25770360
Fax: 22059262

Customer Assistance: