

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة اتكو التكنولوجيه

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهره, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58335

Customer	Alorf Hospital			Equipment	Autoclave.				
				Model No.	Stericlave 18				
				Serial No.	13X 22079				
Authorized Person				Manufacturer	Cominox				
Customer Complaint				Location	1st floor.				
Bovidic Test Error				Contract / Order No.					
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐				☐ Installation			☐ Warranty		
				☒ Repair			☐ On Charge		
				☐ Maintenance			☐ Contract		
				☐ Upgrade			☐		
				Time Spent					
Remarks Service call attended. when we inspect the unit the user says that the bovidic test done on morning not passed. we checked the unit and done the basic tests. Everything is in the limits. we kept the unit for one day observation.									
No.	Spare Parts Used			Part No.		Qty.			
Service Engineer			Authorized Person		Customer / User				
Ni 3com									
ATCO Technology Co.									
Signature			Signature		Signature				
Date 21/02/2024			Date		Date				