

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2195

شركة اتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٢٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 58307

Customer	Alor Hospital		
Equipment	Electro cautery machine.		
Model No.	208 SURGILEC MC 3		
Serial No.	208		
Authorized Person	Manufacturer SURGILEC		
Customer Complaint	Location OR		
Neutral electrode Post pin not connected with the unit.	Contract / Order No.		
Services / Check Carried out	☐ Installation ☐ Warranty		
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks	Service call attended. we inspect the unit and fixed the neutral connection Post. we properly connected the Post with unit and now the unit is working. checked and confirmed That the unit is working properly.		
No.	Spare Parts Used	Part No.	Qty.
Service Engineer	Authorized Person	Customer / User	
Nisam	23/1/24		
ATCO Technology Co.			
Signature	Signature	Signature	
Date 18/01/2024	Date	Date	