

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no-1735

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجاهراء ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63001

Customer	Aleat Hospital Jahra	Equipment	Centrifuge
		Model No.	Z2060
		Serial No.	60130506
Authorized Person		Manufacturer	Hemile
Customer Complaint		Location	LABORATORY
	Aid not opening	Contract / Order No.	

Services / Check Carried out

- | | |
|---|---|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ _____ |

- | | |
|---|---------------------------------|
| ☐ Installation | ☐ Warranty |
| ☒ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ _____ |

Time Spent

Remarks

Some call attended.
Checked the unit and
found the damaged lid lock.
Need replacement for proper
working.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

ATCO Technology Co.

Signature

Date

Santh
28/9/24

Authorized Person

Signature

Date

Customer / User

Hayat Sami
29-9-2024

Signature

Date

د. آيات سامي محمد عبد الله
طبيب مسجل باثولوجي
مستشفى الهرف
22023120560