

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



log no - 2591, 2592

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجھراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 60260

Customer	Alorif Hospital	Equipment	Autoclave
		Model No.	Selectomat PL
		Serial No.	220823 & 220824
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	CSSD
		Contract / Order No.	

Services / Check Carried out

- | | |
|---|---|
| ☒ Inspection / Cleaning | ☐ Optics Checks |
| ☒ Calibration / Performance | ☒ Pneumatic Checks |
| ☒ Interface / Communication | ☐ Hydraulic Checks |
| ☒ Safety Checks | ☒ Operational Error |
| ☒ Electrical / Electronics | ☐ Mechanical Checks |
| ☒ Accessories Check | ☐ |

- | | |
|---|------------------------------------|
| ☐ Installation | ☐ Warranty |
| ☐ Repair | ☐ On Charge |
| ☒ Maintenance | ☐ Contract |
| ☐ Upgrade | ☐ |

Time Spent

Remarks

Preventive maintenance and annual maintenance kit replacement performed for the two units.

Checked and confirmed that the two units are working in good condition.

No.	Spare Parts Used	Part No.	Qty.

under warranty

Service Engineer

Niswan
ATCO Technology Co.

Signature

Date 08/04/2024

Authorized Person

Signature

Date

Customer / User

Signature

Date