

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



Log no - 2591

شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61579

Customer	Abul Hospital	Equipment	CSSD Autoclave.
		Model No.	Seletemat PL
		Serial No.	220823
Authorized Person		Manufacturer	mmm
Customer Complaint		Location	CSSD
		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error		
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____	Time Spent	

Remarks Autoclave unit not chamber cleaning. cleaned the chamber and rack. cleaned the RO water tank and refill with fresh new water. tested one normal cycle and confirm the unit is working properly. we kept unit under observation for one week. if any stain settlement. If again the stain occurs

No.	Spare Parts Used	Part No.	Qty.
	Need to test the water softer water quality.		
	under warranty		

Service Engineer ATCO Technology Co. Signature Date 29/07/2024	Authorized Person Signature Date 1/8/24	Customer / User Signature Date
---	---	--------------------------------------